Optimization of the Heat Bonding Process for Pacemaker Leads: Conceptual Design and Finite Element Analysis

Author: Wilyim Rodriguez Vargas
Advisor: Dr. Hugo Pelaez Carpio
Department: Mechanical Engineering

Abstract

This project aims to optimize the heat bonding process [1] used in manufacturing pacemaker leads by redesigning the thermal assembly to enhance temperature uniformity, process efficiency, and alignment precision. The original system, as shown in Figure 2, which used a dual-stage heating method and INVAR 36 [4] bars with copper inserts, showed substantial thermal gradients and mechanical misalignment. Figure 1 illustrates a redesigned configuration incorporating Celazole [5] thermal insulators was proposed and simulated through Finite Element Analysis [2]. The results indicated a substantial improvement in thermal uniformity and faster stabilization time, reducing peak INVAR temperatures and increasing production reliability. A functional prototype was developed and successfully integrated into the automated manufacturing process.

Introduction

Pacemakers are vital implantable devices that manage abnormal heart rhythms. Central to their operation are **pacemaker leads**, which transmit electrical impulses from the pulse generator to the heart. The structural integrity and electrical continuity of these leads depend heavily on the **quality of the bond** between the polyurethane tube and metallic electrodes, often platinum-iridium. In current manufacturing lines, **heat bonding is performed in a dual-stage process**, but performance inconsistencies have been linked to thermal gradients and material mismatch. The overarching aim of this project is to address these **thermal and mechanical shortcomings** through redesign and simulation-driven validation, ensuring **repeatable, biocompatible, and high-precision assembly**.

Background

The original bonding setup consisted of an INVAR bar with copper inserts and dual-stage heating (320°F pre-heat and 470°F final heat). However, simulations revealed large temperature spreads and mechanical instability, caused by copper's high thermal conductivity and expansion. The literature supported the use of high-performance insulators like Celazole to limit heat propagation.

Problem

Problem Statement
Issues Identified:
•Inconsistent bonding due to **thermal gradient >60°F** across bonding zones.
•Misalignment caused by **thermal expansion of copper inserts** within a rigid INVAR structure.
•Long thermal cycle (~370 seconds) reducing production throughput.
Research Objectives:
1.Design a **thermal isolation system** using Celazole interfaces to reduce lateral heat flow.
2.Optimize heating profiles to stay within the thermal window for polyurethane bonding (below 250°C).
3.Implement **FEA simulations** to predict transient and steady-state temperature profiles and ensure design viability.
4.Validate the prototype under **actual process control** systems (PID, vision alignment).

Methodology

The research followed a structured multi-phase approach:
1. Conceptual Design:
•A redesigned assembly using **Celazole rings** between copper and INVAR was proposed.
2. CAD Modeling:
•3D models created using SolidWorks and exported for FEA integration.
3. Finite Element Analysis (FEA):
•**Transient and steady-state simulations** were performed using SolidWorks Simulation.
•Boundary conditions: Ambient convection (38 W/m²·K), no radiation.
4. Prototype Integration:
•New design built and installed in an existing automated fixture.
•Retained **PID-based heating control** and **machine vision alignment** for component placement verification.



Fig. 1. Pacemaker Lead Assembly



Fig. 3.Original Design Transient State Study



Fig. 5. Original Design Steady State Study

Results and Discussion

Results and Discussion
•**Temperature Spread**: Reduced from ~60°F to <5°F between copper inserts.
•**INVAR Max Temp**: Reduced from 340.6°F to 119.9°F, as illustrated in Figures 4 and 5, respectively.
•**Stabilization Time**: Reduced from 370 seconds to ~80 seconds, as presented in Figures 3 and 4, respectively.
•**Mechanical Stability**: Misalignment issues eliminated.
•**Automation**: PID thermal control [3]and machine vision integration remained fully functional.
These results validate the thermal and mechanical advantages of the redesigned system.



Fig.2. Original Design



Fig. 4. New Design Transient State Study



Fig.6. New Design Steady State Study

Conclusions

This project validates a novel thermal bonding system that resolves the core limitations of traditional pacemaker lead manufacturing. By introducing **Celazole thermal insulators**, the assembly achieves:
•Improved **temperature uniformity**
•Dramatically **reduced stabilization time**
•Maintained **mechanical alignment**
•Seamless **integration with existing automation**
The success of the simulation-driven approach confirms that **high-precision medical device manufacturing** can be enhanced through robust thermal design and materials engineering.

Future Work

1. **Experimental Bond Strength Testing**: Pull tests to quantify bonding integrity across thermal cycles.
2. **Thermal Imaging Validation**: Use IR thermography to validate simulation accuracy.
3. **Advanced FEA**: Include **radiation modeling and variable convection** to capture more realistic environments.
4. **Geometric Optimization**: Explore insert designs that further minimize lateral conduction.
5. **Technology Transfer**: Apply methodology to other thermally bonded medical assemblies (e.g., catheter shafts, neurostimulation leads).

Acknowledgements

I would like to thank **Dr. Peláez Carpio** for his invaluable guidance and mentorship throughout this project. Appreciation is also extended to the **Mechanical Engineering Department** at the **Polytechnic University of Puerto Rico** for providing the resources and support required for the design, simulation, and prototype testing phases.

References

[1] C. Ageorges and L. Ye, “Fusion Bonding of Polymer Composites”. London, U.K.: Springer, 2002.

[2] N. H. Kim, B. V. Sankar, and A. V. Kumar, “Introduction to Finite Element Analysis and Design”, 2nd ed. Hoboken, NJ, USA: Wiley-Blackwell, 2018.

[3] K. Ogata, “Modern Control Engineering”, 5th ed. Upper Saddle River, NJ, USA: Prentice Hall, 2010.

[4] Carpenter Technology Corporation, "Invar 36® alloy data sheet" [PDF]. [Online]. Available: https://carpentertechnology.com/hubfs/7407324/Material%20Saftey%20Data%20Sheets/Invar%2036.pdf. Accessed: Apr. 30, 2025.

[5] PBI Polymer, "Celazole® material information." [Online]. Available: https://pbipolymer.com/products/celazole-pbi. Accessed: Apr. 30, 2025.

[6] Dassault Systèmes, "Performing thermal analysis," SOLIDWORKS Help, 2025. [Online]. Available: https://help.solidworks.com/2025/English/simconntutorial/Performing_Thermal_Analysis.htm. Accessed: Apr. 30, 2025.